

Konformitätserklärung Declaration of Conformity

Wir

We

**B. Braun Melsungen AG
Carl-Braun-Str. 1
34212 Melsungen
Deutschland/Germany
SRN DE-MF-00000201**

erklären in eigener Verantwortung,
dass das/die Produkt/e

**Power cord EU
Power cord GB
Power cord US
Power cord AU
Power cord CH
Power cord ZA, IN
Power cord CN
Power cord DK
Power cord BR
Power cord AR**

**Connection Lead 12V CP
Staff call cable compactplus**
Zubehör für compact^{plus} Infusionspumpen
Basis UDI-DI: 403923900000222ZG
(Artikelnummern siehe Anlage I)

mit den Anforderungen der Medizinprodukte
Verordnung (EU) 2017/745
übereinstimmt/übereinstimmen

Konformitätsbewertungsverfahren
nach Artikel 52 Absatz 7 der oben genannten
Verordnung

Klassifizierung
gemäß Anhang VIII der oben genannten
Verordnung
Klasse I

**Des Weiteren erklären wir in eigener
Verantwortung, dass oben genannte
Medizinprodukte die Anforderung zu
folgender EU Richtlinie**

Richtlinie 2011/65/EU

hereby declare in our own responsibility
that the product/s

**Power cord EU
Power cord GB
Power cord US
Power cord AU
Power cord CH
Power cord ZA, IN
Power cord CN
Power cord DK
Power cord BR
Power cord AR**

**Connection Lead 12V CP
Staff call cable compactplus**
Accessories for compact^{plus} Infusion Pumps
Basic UDI-DI: 403923900000222ZG
(article numbers see attachment I)

is/are in conformity with the requirements of the
Medical Device Regulation (EU) 2017/745

Conformity Assessment Procedure
according to article 52 section 7
of the Regulation named above

Classification
according to annex VIII of the Regulation named
above
Class I

**Furthermore, we declare in our own
responsibility that the above-mentioned
medical devices meet the requirements of
the following EU Directive**

Directive 2011/65/EU

des Europäischen Parlaments
und des Rates vom 8. Juni 2011
zur Beschränkung der Verwendung bestimmter
gefährlicher Stoffe in Elektro- und
Elektronikgeräten

erfüllen

Gültig bis
gemäß gültigem ISO Zertifikat
Q5 012974 0590

of the European Parliament and
of the Council of 8th June 2011
on the restriction of the use of certain
hazardous substances in electrical and
electronic equipment.

Valid until
according to our valid ISO Certificate
Q5 012974 0590

Anlage I / Attachment I**Basic UDI-DI 403923900000222ZG**

Art.-Nr. / Art. No.	Produktname / Product name	Klasse / Class
8717110	Power cord EU	
8717111	Power cord GB	
8717112	Power cord US	
8717113	Power cord AU	
8717114	Power cord CH	
8717115	Power cord ZA, IN	
8717117	Power cord CN	
8717118	Power cord DK	
8717119	Power cord BR	
8717121	Power cord AR	
8718020	Connection Lead 12V CP	
8718030	Staff call cable compact ^{plus}	

Document amendment information

Version	Description of the changes
1.0	Initial version of document.
2.0	Adaption of validity.
3.0	Adaption of validity.

Title: Declaration of Conformity - 170-102 - Accessories for compactplus Infusion Pumps - MDR Initiator: Anna Heil

This document is signed electronically in compliance with the B. Braun electronic signature policies and procedures by following persons:

UserName: Heil, Anna (heilaade)
Title: Manager Regulatory Affairs
Date: Monday, 22 January 2024, 10:52 W. Europe Daylight Time
Meaning: Document signed as Author

UserName: Arico, Mareike (sommrde)
Title: HC-QM - Head of QM active MD/ Head of Regulatory Affairs CoE AIS
Date: Wednesday, 24 January 2024, 12:49 W. Europe Daylight Time
Meaning: Approve Document

UserName: Loh, Malte (lohmatde)
Title: HC-RA-DE08 Senior Manager Regulatory Affairs
Date: Friday, 26 January 2024, 16:30 W. Europe Daylight Time
Meaning: Approve Document
